



Product: **COHSII Media**
0.2µm sterile-filtered

Catalog #: 900-750
Lot #: M94Q00J

Manufacture Date: Jan 2018
Expiration Date: Jan 2020

TEST	METHODOLOGY	SPECIFICATION	RESULTS
Biological Testing			
Endotoxin	USP <85>	Report	<0.500 EU/mL
Microbiological Testing			
Sterility			
Bacteria		Not Detected	Not Detected
Fungi		Not Detected	Not Detected
Mycoplasma	Barile, MF and Kern, J (1971)	Not Detected	Not Detected
General Safety Testing*	USP <88>	Pass	Pass
Physical Testing			
Osmolality		275 - 340	298 mOsm/kg
pH		7.3 - 7.5	7.47

*Note: The sample meets requirements of the General Safety Test under the United States Pharmacopeia and National Formulary, Chapter <88> Biological Reactivity Tests.



Created by



Verified by